

www.dewesoft.com - Copyright © 2000 - 2025 Dewesoft d.o.o., all rights reserved.

Rotational and Torsional Vibration Measurement and Analysis



Rotational and Torsional vibrations

Vibrations are mechanical oscillations, that occur around an equilibrium point. The oscillations can be periodic or random.

Vibrations of any mechanical device in operation are typically unwanted. Such vibrations can be caused by imbalances in the rotating parts, uneven friction, the meshing of gear teeth,... We can minimize unwanted vibrations with proper designs.

We can decompose any vibrating system into elementary parts:

- spring a means of storing potential energy (k)
- mass a means of storing kinetic energy (m)
- damping a means by which energy is gradually lost (c)

Vibrations can be classified as:

- free (after the initial disturbance, the system is left to vibrate on its own) and forced (the system is subjected to an external force, often a repeating type of force)
- damped and undamped
- linear (all the basic components of a vibration system the spring, the mass, and the damper behave linearly) and nonlinear (one or more basic components of a vibration system are not linear)
- deterministic (magnitude of the excitation (force or motion) acting on a vibratory system is known at any given time) and random (the value
 of the excitation at any given time can not be predicted wind velocity, road roughness and ground motion during an earthquake)

When the system vibrates with its own natural frequency, it falls into resonance - it oscillates with the largest amplitude at the same excitation force.

Quantities that are used to describe the vibrations are:

- displacement (in [m])
- velocity (in [m/s])
- acceleration (in [m/s2])

All three quantities are directly related to each other - velocity is a derivation of displacement and acceleration is a derivation of velocity.

Torsional vibration

Torsional vibrations are angular vibrations of an object, typically a shaft along its axis of rotation. Torsional vibrations are evaluated as the variation of rotational speed within a rotation cycle. RPM variations are typically induced by a rough driving torque or a varying load.



load, RPM, etc.).

Torsional vibrations are important whenever power needs to be transmitted using a rotating shaft or couplings, such as in the case of automotive, truck and bus drivelines, recreation vehicles, marine drivelines or power-generation turbines.

Quantities that are used to describe the torsional vibrations are:

- angular position (in [rad] or [°])
- angular velocity (in [rad/s] or [°/s] or [RPM])
- angular acceleration (in [rad/s2] or [°/s2])

All three quantities are directly linked to each other and can be derived by integration and differentiation. The most common quantity used to measure torsional vibrations is the angular velocity or RPM.

The source of excitation is quantified by torque.

The angular velocity and displacement quantify the response of the applied torque. Torque and angular velocity are linked to each other:

$$T = \frac{d^2 \alpha}{dt^2} J_z$$

Where T is torque in [Nm] and Jz is the torsion constant or the polar moment of inertia in [m4].

This formulation is similar to the formulation of Newton's law, which links the force acting on a mass to the acceleration of that mass.

Torsional vibrations in Dewesoft X

The torsional vibration software option of <u>Dewesoft</u> is used to obtain a rotational/torsional vibration monitoring and analyzing solution, for research, development, and optimization. With the small form factor of the Dewesoft instruments (e.g. <u>SIRIUS</u>, <u>DEWE-43</u>) the perfect mobile solution for test engineers and consultants is born.



Torsional and rotational vibration is calculated, as well as the corresponding velocities. The software can compensate for off-center mounting of the sensor and can also take care of the gearbox ratio. Furthermore, the possibilities of <u>Dewesoft</u> allow angle-domain visualization and export.

Using <u>Dewesoft</u> to capture data from other sources: e.g. Video, CAN, Ethernet means that they are perfectly synchronized in one data file. If the powerful integrated post-processing features of <u>Dewesoft</u> are not enough, you can even export the data to several different file formats.

In addition to torsional vibration, the system can be expanded with the order tracking option to complete the picture of the measurement situation.

System overview

As the torsional vibration measurements are very critical in terms of accuracy, only precise counter sensors are supported. At least one (rotational vibration), but usually two (torsional vibration) encoders are necessary.



Image 3: An example of an angle sensor - encoder

- Encoder
- CDM with zero pulse (Geartooth)
- Zebra tape sensor (with DS-TACHO4 and auto-gap detection)
- Sensor with missing teeth (e.g. 60-2)

For example, The CA-RIE sensor has a resolution down to 0.5Ű and is, therefore, less sensitive to vibrations (which could damage standard encoders over time).



Image 4: Supported angle sensors

Enabling Torsional vibration module

Like many additional mathematical modules, the Torsional vibration is an option to the standard Dewesoft package and needs to be enabled in the Math section.

	Q	Add module New setup defaults			
General		👆 User inputs		Electrical measurements	
Channels		Cursor		F Power analysis	×
Storing		Frequency domain analysis		Acoustics	
Data header	*	III FFT analysis))) 🕅 Sound level meter	
System monitor		Machinery diagnostics		Aerospace and automotive	
	*	Combustion engine analysis	メ	Navigation	
M Analog in	*	(🖴) Human vibration		Psophometer	
Function generator	*	🔨 Modal test		()) Psophometer	
CAN 010001	*	Order tracking	*	Vibration	
Counters	*	(Rotor balancer	×	Sine processing	
M Analog/digital out	*	Historian Torsional vibration		SRS	
*	م د			The one	

Torsional vibration 1.0.0

Torsional vibration application module allows measure dynamic and static bending and vibration of the shafts. The torsional vibration also measures the twist of the shaft with higher RPM. With torsional vibration module two different parameters can be measured: rotational vibration and torsional vibration.

Image 5: Adding Torsional vibration module in Dewesoft

Basic operating concept

The Torsional vibration module inside <u>Dewesoft</u> is just one of several other application modules that offer dedicated mathematics and dedicated visual controls like angle based XY-diagram.



Image 6: Basic operating concept

General setup

The user interface of the Torsional Vibration module is split into the following sections:

TV 1 +		Used V Setup V Rename
Input	Input channel configuration	
First sensor input CNT1 V	Sensor 1	Sensor 2
Second sensor input CNT2	Encoder-512 V 1 / 1 Encoder-512, X1	Encoder-360 V Encoder-380, X1
<- Sensor_1_angle ->	Filter settings	
Output	Input filter Rotational DC filter	
Name Sensor_1_angle	off v 10 v Hz	
Description -	Sensor 1 output channel Torsional vibration channel	els Sensor 2 output channel
Units deg Color	Rotational angle Torsional angle Rotational velocity Torsional velocity	Rotational angle Rotational velocity
Max value 360 deg	✓ X axis reference angle ✓ Frequency [RPM]	
Max 36,4 deg RHS 0,4841 deg Average 35,59 deg Min 34,76 deg	Angle offset	teference curve Set Use reference curve Resolution not set
Min value 0 deg		(Gep) előu
+ -	0,00 50,00 100,00 150,00 X axis (200,00 250,00 300,00 360,00 ₹

Image 7: Channel setup of Torsional vibration math module

- input: Define the counter channels for input (e.g. CNT 7, CNT 8).
- counter setup: Define a type of angle sensor (e.g. Encoder-512, CDM-360, ...), and select ratio, if gearing is used.
- filter setup: The input filter is needed to prevent glitches and spikes in the signal. The Rotational DC-filter needs to be set to cut the DC component out of the RPMs. We need to set the filter to include all wanted frequencies, but not too low, else we will have static DC deviations on the output signal.
- output channels: Select which calculations are to be done (e.g. rotational velocity, torsional angle....).
- angle offset: Compensate constant angle offset, off-center mounting, sensor errors.
- output: Preview of output channels switch through with the arrow buttons.
- preview: Preview of torsional angle.

Switch to View Channel List on the right upper corner to see a list of the calculated output channels.

TV	1 +				Us	sed	~ 0	Channel list	~	Rename	
Sear	ch		Q								
+	Used	С	Name	Min	Value		Max	Unit			
4	Used		Torsional vibration		Delta_fi, Delta_dfi/dt						
			Sensor_1_angle	0,00	41,50	(deg)	360,00	deg			
			Torsion_velocity	-5,00	-50,1843 / -26,5320 (d	eg/s)	5,00	deg/s			
			Torsion_angle	-5,00	-25,6492 / -22,1036	(deg)	5,00	deg			
			TV_Frequency	0,00	2,2 (RPM)	6000,00	RPM			
			RotAngle_1	-5,00	-0,0097	(deg)	5,00	deg			
			RotAngle_2	-5,00	-0,3353 / 0,3724	(deg)	5,00	deg			
			RotVelocity_1	0,00	0,0 (RPM)	6000,00	RPM			
			RotVelocity_2	0,00	-0,4 (RPM)	6000,00	RPM			
			Rev. count	0,00	0		6000,00				

Image 8: View channel list

Counter sensor editor

There are some typical sensors predefined. However, if your type is not listed, you can define your own sensor in the Counter sensor editor. For torsional vibration module you can use different types of sensors:

- Encoder
- CDM with zero pulse (Geartooth)
- Zebra tape sensor (with DS-TACHO4 and auto-gap detection)
- Sensor with missing teeth (e.g. 60-2)

Go to Options Editors Counter sensor editor:



Image 9: Path to Counter sensor editor

Click on Add sensor first, then enter a name, and set all other parameters. Depending on the sensor different parameters are available. When finished, click Save&Exit.

Current sensor		
My Encoder 1000		
Sensor type	Signal level	
Encoder V	Signal type Signal filter Digital (TTL level) V	
Encoder setup	Signal edge	
Pulses per revolution Default encoder mode 1000 X1	Positive ~ Zero pulse edge Positive ~	

Image 10: Add your own sensor

Now the sensor can be accessed from the dropdown menu of the torsional vibration module.

Input channel configuration	
Sensor 1	Gearbox ratio
Encoder-512 V	· 1 / 1
Encoder-360 Encoder-512 Encoder-900 Encoder-1024 Encoder-1800 Encoder-3600 CDM-360	al DC filter
5. CDM-720	Torsional vibration char
Geartooth proximity Encoder-256 Tape consor My Encoder 1000	Torsional angle Torsional velocity
Frequency [RPM]	

Image 11: Manually defined sensor can be selected in

Difference between rotational and torsional vibration

It is important to know that we actually have two different parameters that can be measured with the torsional vibration module: rotational vibration and torsional vibration.

Rotational vibration is simply the dynamic component of the rotational speed. If we measure the rotational speed of the shaft with high precision, we will notice that we get a high deviation of rotational speed in some regions of the run-up. This is caused by the angular vibration crossing the angular natural frequency of the shaft. It is calculated by cutting off the DC component of the rotational speed or the rotation angle. In the graph below we see two curves, the one above is the RPM, and the curve below shows the deviation in degrees, which is actually the rotation angle vibration. During a coast down the max. vibrations appear at the same RPMs again.

The graph below shows the run-up and coast down where we can nicely see the static twist of the shaft, and when passing through the natural frequency, the angular vibration of the shaft reaches its maximum.



Torsional vibration is an oscillation of angular motions (twist) which occurs along with rotating parts, such as gear trains, crankshafts or clutches. We need two encoders to measure the torsional vibration, so the torsional vibration is actually a difference between angles of the two encoders. The torsional vibration also measures the static twist of the shaft with higher RPMs.



 Required hardware
 SIRIUS ACC+, MULTI or DEWE-43

 Required software
 Dewesoft X, SE or higher + Torsional vibration option, DSA or EE

 Setup sample rate
 At least 10 kHz

Torsional and rotational vibration can be measured with either an encoder (up to 3600 pulses per revolution) or a special sensor (CA-RIE-360/720) that has a lower resolution (up to 720 pulses per revolution) but is much less sensitive to vibrations that could damage standard encoders.



Image 14: Angle sensor CA-RIE

Rotational vibration setup

For our test, we take an electric motor (in the middle) and two encoders. The encoder on the left is connected with coupling and the encoder on the right with a spring to the motor, to create high vibrations.

Because we are currently only interested in rotational vibration, we will only use the encoder on the right side, connected with the spring. The encoder is connected to the Dewesoft instrument on a Counter input.



Image 15: Rotational vibration demo device

As mentioned, there is nothing to be set up in the analog or counter channels setup. Just go to the Torsional vibration module. Next, select the First sensor input. If we have connected the first sensor to CNT 7, we need to select it from the list. Then we define the sensor. If we have 1800 pulses per revolution, we choose the Encoder-1800. If the sensor used is not defined so far, we need to create it in the Counter sensor editor first.

Next, we need to set the Input filter for the counters. The input filter is needed to prevent glitches and spikes in the digital encoder pulse signal. It can be set from 100ns to 5us, the optimal setting is derived from the following equation:

$$InputFilter[s] \leq rac{1}{10 \cdot rac{RPM_{max}}{60} \cdot PulsesPerRevolution}$$

Example: Let's say that our machine is running at 3000 RPM and we have the encoder with 512 pulses per revolution. If we insert this number in the equation above, we get the next result:

$$InputFilter[s] \leq rac{1}{10 \cdot rac{RPM_{max}}{60} \cdot PulsesPerRevolution} = rac{1}{10 \cdot rac{3000}{60} \cdot 512} = 3,9 \mu s$$

The rotational DC-filter needs to be set to cut the DC component of the RPMs. We need to set the filter to include all wanted frequencies, but not too low, or else we will have static DC deviations on the output signal. It can be set from 0.1 to 10 Hz. You have to ensure that your lowest RPM is not filtered out!

A 10 Hz filter, for example, would mean, frequencies below 600 RPM would be suppressed.

Output channels

TV 1 +			
Input			Input channel configuration
First sensor input			
CNT1		~	Sensor 1
Second sensor input			
<no channel=""></no>		\sim	Encoder-1800 ~
			Encoder-1800, X1
<- Sensor	_1_angle	->	Filter settings
Output			Input filter Rotational DC filter
Name Sensor_1_	angle		off ~ 10 ~ Hz
Description -			Sensor 1 output channel
Units deg	Color		Rotational angle
Preview Values		_	Rotational velocity
Maxwalua	260 dag		Frequency [RPM]
riax value	500 deg		
Мах	172,4 deg		
RMS	0,2038 deg		
Min	172 deg	_	
Min value	0 deg		

Image 16: Rotational vibration setup

The output channels are:

- Rotational angle: Filtered angle value of vibration.
- Rotational velocity: Filtered velocity vibration value.
- X-axis reference angle: The reference angle, which is always from 0 to 360 and can be used as a reference for angle based xy diagrams.
- Frequency: In RPM units.

Rotational vibration measurement

When you switch to measure mode, the calculated channels of the Torsional vibration module are shown in the channel selector on the right side. The first idea would be to select the TV_Frequency channel and display in analog/digital meter or a recorder.



Image 17: Run-up can be displayed on a recorder widget

The Sensor_1_angle is the reference angle and can be used for an angle-based display of XY recorder. Add an XY recorder and click on Sensor_1_angle first (=x axis), then on RotAngle_1 (=y axis). Then set the recorder to Angle-based x-y in the properties on the left side. Select e.g. 2 periods to be displayed. This XY recorder now displays the rotational angle of the current revolution. It is like a scope, but with an angle reference instead of a time reference.

Vary the RPM and when you come close to the resonance frequency, the amplitude reaches its maximum.



Image 18: Rotational angle close to the resonance frequency reaches maximum

In the next graph also the rotational velocity is added (first derivation of the angle), according to the theory it is shifted 90 degrees.



Image 19: Rotational angle and rotational velocity are shifted for 90 degrees

Torsional vibration setup

The next step is to measure the torsional vibration.

On the left side is an encoder with a rigid (coupled) connection to the motor, on the right side is another encoder. Both encoders are connected with Dewesoft hardware in counter channels. We measure the torsion vibrations of the metal spring in between.



Image 20: Torsional vibration demo equipment

Add a Torsional vibration module. We need to select both input channels in the Torsional vibration setup. We select CNT0 as the first channel and CNT1 as the second channel. Then we need to define Sensor 1 and Sensor 2 by selecting the correct type from the dropdown menus on the right side. If we have a gearbox in between, we need to enter the Gearbox ratio (it is used for measuring the torsion angle across the gearbox).

As already explained in the rotational vibration section, you can use the input filter for cleaning the signal from eventual glitches, and the rotational filter removes the DC offset from the difference signal.

TV 1 +			Used V Setup	Rename
Input	Input channel configuration			
First sensor input CNT1 V	Sensor 1	Sensor 2	1	
Second sensor input CNT2	Encoder-512 V	Searbox ratio 1 / 1 Encoder-1800 Encoder-1800, X1	~	
<- Sensor_1_angle ->	Filter settings			
Output Name Sensor_1_angle	Input filter Rotational DC filter			
Description -	Sensor 1 output channel	Torsional vibration channels	Sensor 2 output channel	
Units deg Color Preview Values	Rotational angle Rotational velocity X axis reference angle	Torsional angle Torsional velocity	Rotational angle Rotational velocity	
Max value 360 deg	Frequency [RPM]			
Max 357,1 deg RMS 1,006 deg	Angle offset Zero Reset zero Use for Torsional angle with	Reference curve	Use reference curve Resolution not set	
Average 355,2 deg Min 353,6 deg		Torsion angle		
Min value 0 deg	Torsion_angle (deg)			
Templates Save	다. 다. - 다. 다. 0,00 50.0010	0.00 150.00 <u>200.00</u>	250,00 300,0	0 00000 Angle (deg)
T		X axis ()		

Image 21: Torsional vibration setup

The Input filter can be set in a range between 100ns and 5ms. The optimal settings are derived from the following equation:

$$InputFilter[s] \leq rac{1}{10 \cdot rac{RPM_{max}}{e0} \cdot PulsesPerRevolution}$$

We have the following Output channels:

- Torsional angle: the Dynamic torsional angle is the angle difference from sensor 1 to sensor 2.
- Torsional velocity: Difference in angular velocity from sensor 1 to sensor 2.
- Sensor 1 Rotational angle.
- Sensor 2 Rotational angle.
- Sensor 1 Rotational velocity.
- Sensor 2 Rotational velocity.
- X-axis reference angle: The reference angle, which is always from 0 to 360 and can be used as a reference in angle based xy diagrams.
- Frequency: In RPM unit.

Angle offset and Reference curve configuration

In the section Angle offset, you see the angular difference between the two sensors. Click on Zero to remove this static offset. The current average value of the signal will be subtracted. Then click on the y-axis to auto-scale the signal.



Image 22: Angle offset section in Torsional vibration setup

Reference curve configuration

Now there is also an option on how to compensate for off-center mounting and unsteady pulses from the encoder.

Centered mounting is very important. In the first picture (1) there is no problem, the disk is mounted perfectly. However, in real life, off-center mounting like in the second picture (2) will appear. Let's draw on the disk a red line on 0 degrees and a blue line on a 90-degree position. When the disk is turned, the sensor (black box) will count the pulses and after a certain number detect the 90-degree position. But if you look at the disk now, it is far from 90 degrees!



Image 23: Off-center mounting can cause issues

Imagine it rotating, then a constant sine wave will be generated additionally to the rotational vibrations. This can be compensated in <u>Dewesoft</u> using the **Reference curve**. However, it is required that the load must be removed from the engine! It must be free-running. Otherwise, you would also cancel out vibrations you want to analyze. When the machine is running at idle speed with no real torsional vibration, press the **Set** button.



Image 24: Setting the reference curve

The current curve is now recorded over one revolution as a reference. After that, the line is much flatter.

Torsional vibration measurement example

Now let's do the measurement. We have added the analog and the digital meter for frequency, a recorder for the frequency, the torsional angle and torsional velocity and we have also added angle based XY for those parameters.

The run-up and run-down clearly shows the static torsion angle and the region of the natural frequency with high torsional vibration values.



Image 25: Torsional vibration measurement screen

Order extraction

To extract orders from a torsional/rotational vibration, we need to add the order tracking module. For the frequency source, we need to define the Torsional vibration module. For this, you select the module created before and Frequency channel(within the module if torsional vibration is used). Next, we need to define the Upper RPM limit and Lower RPM limit. This is used to reserve the memory for waterfall FFT. The waterfall will be drawn from the lower to the upper limit in the Delta RPM step. In this case, we will have (3000-0)/50 = 60 steps of a waterfall.

We choose to extract the first three orders by entering 1;2;3 in the Harmonics field.

Frequency channel setup					
Frequency source N	Module	Frequency channel			
Torsional vibration \checkmark	Torsional vibration 1 \sim	CNT 0 ~			
Calculation criteria - frequency limit	5				
Direction RPM	Upper RPM limit	1			
Delta RPM	Lower RPM limit	7			
50					
Order FFT setup	Common properties		Time FFT setup		
Order domain harmonics	Harmonic list 1; 2;	3	Time domain harmonics		
🗹 Order waterfall vs. RPM	FFT window Black	nan V	FFT waterfall vs. RPM		
Order resolution $$ 1/8 (0.125) $$ $$ $$	Amplitude RMS	~	Auto 🗸 (df = 4,883 Hz)		
1 A	Data averaging Avera	ge between classes 🛛 🗸	update on RPM change (Delta RPM)		
	Update criteria Alway	s v	update on 0 min 0,5 s		
1 2 3 ord	RPM		Overall RMS vs. RPM		
Maximum order 32 ~			(rms calculation depends on FFT waterfall resolution)		
Order waterfall vs. time		→ t			
Overlap 0 v %	Skip missing RPM class				

Image 26: Order extraction using Order tracking module

Now let's do some measurements. We have frequency and harmonic channels, therefore we can add the x-y display for displaying the orders. The OT_Frequency should be used as x-reference (first chosen channel of x-y), and the RotAngle_1/Amplitude1, RotAngle_1/Amplitude2 and RotAngle_1/Amplitude3 orders should be used as the y-axis. Then we make a run down to observe the orders, where we can clearly see the peak at 2422 RPMs that we previously observed in the recorder being the first order of vibrations.



Image 27: Observing the behavior of orders

To see it even better, we can add the 3D graph. This graph can only show matrix channels. Order tracking has two of them - Order waterfall and FFT waterfall, which can be used as data sources. Now we can start our run up or run down. Since the color shows the amplitude at a certain frequency and RPMs, we need to carefully choose the amplitude scale to show the graph in a nice way. To see also the small amplitude values, the logarithmic y-scale is recommended.



Furthermore, the amplitudes/phases (or even real and imaginary part) of the Orders can be plotted against the frequency. Again, take the XY recorder, but now set graph type to Single x-axis to get a Bode plot. Here you see the first harmonic (green line) with its maximum around 2400 RPM, matching nicely to the previous measurements. The turning of the phase also indicates resonance.



FFT peak calculation

One of the standard measurements is, to do the run-up of the machine, and then calculate the max amplitude over the FFT.

Add an FFT analysis function from the Frequency domain analysis section in math.

।	Add module New setup defaults		
General	👆 User inputs	Electrical measurements	
Channels	Cursor	🗲 Power analysis 🗙 🗙	
Storing	Frequency domain analysis	Acoustics	
🛢 Data header 🗙 🗙	III FFT analysis)) ⑦ Sound level meter	
· · · ·			
FFT analysis 1.0.0 Fourier transform application is a transformation (used to transform signals between time domain an	d frequency domain. Output of the transform could be	

rourier transform application is a transformation used to transform signals between time domain and frequency domain. Output of the transform could be complex, amplitude or overall rms. Calculation types are Overall (Averaged) and Block history. Window type and resolution of the FFT are calculation parameters

Image 31: Adding FFT analysis module

Then select the input channel, here for example Torsion_angle. Set to Amplitude, Overall, and averaging type Peak:

	Output spectrums	Amplitude scaling	Frequency Weigh	iting Output axis unit	ŧ
	🗹 Amplitude	Peak ~	Lin (Z) 🗸 🗸	Hz	-
I	Complex				
	Output scalars	Scalar outputs fre	quency range		
	Rms	Start frequency	End frequency		
	Peak 🗌	0	- 2497,5586	Hz	
	Peak-Peak	Full frequency	range		
	Calculation type				
	O Block history			Averaging	type
	• Overall (Averaged)			Peak	~
	Stop calculation				
	Stop after	4	blocks \sim		
	Resolution settings				
	Resolution				
	Lines ~	1024 ~	(lines = 1024, df =	2,44 Hz, duration = 0,4	1 s)
il					
1	Window functions				
	Window				
	Blackman 🗸				
	DC cutoff	Overlap			
	None v Hz	0 ~	%		

Image 32: FFT analysis setup

Here is an example, done offline on a data file (of course you also can do it online, during measurement). A section has been selected in the recorder instrument below (green line) to only analyze the run-up of the machine.

Then a 2D graph was added (see instrument bar, red box) and the AmpIFFT math channel assigned.



Image 33: Displaying FFT peak results on a 2D graph

You can do the standard export, or just click on the 2D instrument, and use Options Copy to clipboard "Widget data".



Image 34: Copy the widget data to the clipboard

Pasting, e.g. into Excel gives the following result:

	А	В
1	Freq [Hz]	FFT analysis 1/Torsion_angle/AmpIFFT [deg (Z)]
2	0	2,4606
3	9,77	2,9429
4	19,53	0,8556
5	29,3	3,4773
6	39,06	6,4729
7	48,83	4,3536
8	58,59	0,8813
9	68,36	0,133
10	78,13	0,1083
11	87,89	0,096
12	97,66	0,0948
13	107,42	0,06
14	117,19	0,0335
15	126,95	0,0298
16	136,72	0,0244
17	146,48	0,0225
18	156,25	0,0208
19	166,02	0,0193
20	175,78	0,0184
21	185,55	0,0181

Image 35: Pasting the widget data into Excel

Angle based data export

The angle-based display of measurement data is often very useful. In the XY recorder you assign the reference angle Sensor_1_angle to the x-axis; then set the view either to:

• Single x-axis: sees all cycles layered over each other (persistence mode), easy to see min/max overall cycles; persistence can be adjusted with the Pretime limit in the properties.



Image 36: Single X-axis XY recorder

• Angle-based X-Y: we can see the current cycle - it is working like a scope.



Image 37: Angle based XY recorder

Now, how is it possible to export this cycle-based data? Go to Math and add a Time-to-vector transform.

٩	Add math Manage favorites		
Formula and scripting	Time domain analysis		
C↔→ C++ script	-5 ms Delay channel		
X²+Y² Formula	Tatch value math		
Filtering	Cope math		
FIR filter	∫(dt) Time integration, derivation		
Frequency domain filter	Time-to-vector transform		
TIR filter	Frequency domain analysis		
Image 38: Adding new math Time-to-vector transform			

With this great tool, <u>Dewesoft</u> arranges the input data, e.g. Torsion_angle over the Sensor_1_angle in a matrix by resampling the values to the angle base, you also can select the resolution, in this example, we just use 1 degree. The output data is Actual, Average, Minimum, and Maximum over all the cycles.

Input	Output channels		
tor 🛞	Actual	Minimum	
Torsional angle	Average	Maximum	
	Settings		
	Reference ch.	Sensor_1_angle	~
		from	to
	Range	0 Grad	360 Grad
	Step	1 Grad	
<- Torsional angle/AVE ->			
Output			
Name Torsional angle/AVE			
Description -			
Units N Color			
Preview Values X axis			
- (N)			
50,000-			
-50,000-0,000 180,000 380,000			

Image 39: Time-to-vector transform setup

For displaying the matrix data, you have to use a different visual instrument, the 2D graph. Then you can assign the ACT (actual), MIN (minimum), MAX (maximum), or AVE (average) channels from the Math function.



Image 40: Displaying the results from Time-to-vector transform on a 2D graph

Store a datafile, afterward in the Export section only select the 3 matrix channels (in order that data has the same format):

12	No	Basic stat. 3 (Basic statistics)	Kraft/MAX 2/MAX	Asynchronous	1,6 Hz	Scalar
13	Yes	T-V transf. 1 (Time-to-vector transf	Torsional angle/AVE	Single value	unknown	Vector (361)
14	Yes	T-V transf. 1 (Time-to-vector transf	Torsional angle/MIN	Single value	unknown	Vector (361)
15	Yes	T-V transf. 1 (Time-to-vector transf	Torsional angle/MAX	Single value	unknown	Vector (361)
16	Yes	T-V transf. 1 (Time-to-vector transf	Torsional angle/ACT	Single value	unknown	Vector (361)

Image 41: Export the data from Time-to-vector transform math

Here we have done the export to Excel, instead of the time axis, we have angle-based data.

	А	В	С	D	E	F	G	н	J	MV	MW	MX	MY
1	Torsional angle	deg	0	1	2	3	4		356	357	358	359	360
2	Torsional angle/AVE	deg	-8,63393	-8,72163	-8,79453	-8,67877	-8,63295	-8	,28899	-9,22222	-9,17146	-8,90385	-8,88429
3	Torsional angle/MIN	deg	-12,0519	-12,5902	-12,8269	-13,9257	-11,5307		13,4678	-12,9858	-12,6452	-12,7035	-13,3138
4	Torsional angle/MAX	deg	-7,27313	-6,65147	-7,02768	-5,83716	-6,9098		-7,41747	-7,33984	-7,28405	-7,15611	-7,09349
5	Torsional angle/ACT	deg	-8,91775	-7,74995	-8,33797	-8,96986	-10,3347		-8,19921	-10,3526	-11,7044	-9,10835	-7,95053

Image 42: Angle-based data in Excel

If you copy the data, then use the paste special... function in Excel, you can Transpose columns and rows automatically and get the angle from topdown instead of left to right.

Paste Special	? ×	4	А	В	с	D	E
Paste		1	Torsional	Torsional angle/AVE	Torsional angle/MIN	Torsional angle/MAX	Torsional angle/ACT
All	○ All using Source theme	2	deg	deg	deg	deg	deg
O <u>F</u> ormulas	All except borders	3	0	-8,6339273	-12,051939	-7,2731271	-8,9177542
○ Values	○ Column <u>w</u> idths	4	1	-8,721633	-12,590186	-6,6514735	-7,7499509
○ Forma <u>t</u> s	○ Fo <u>r</u> mulas and number formats	5	2	-8,7945347	-12,8269	-7,0276752	-8,3379707
○ <u>C</u> omments	○ Val <u>u</u> es and number formats	6	3	-8,6787663	-13,925694	-5,8371649	-8,9698648
🔿 Validatio <u>n</u>	○ All merging conditional formats	7	4	-8,6329508	-11,530666	-6,9098001	-10,334683
Operation		8	5	-8,3981962	-12,069989	-6,5598674	-8,6783314
● N <u>o</u> ne	O Multiply	9	6	-8,0143414	-12,286983	-6,2056117	-6,4751434
○ A <u>d</u> d	O D <u>i</u> vide	10	7	-7,3506231	-11,375021	-5,2072558	-6,1924405
○ <u>S</u> ubtract		11	8	-6,6269274	-10,612842	-4,8699379	-5,235064
		12	9	-5,8668566	-9,4796171	-4,2662416	-5,1964402
Skip <u>b</u> lanks	Transpose	13	10	-5,4541864	-9,9645853	-3,6833789	-5,0723882
Parte Link	OK Cancel	14	11	-4,8030043	-7,569243	-2,8832061	-4,0516591
Paste LIIIK	OK Cancel	15	12	-4,1257186	-7,0892282	-2,7019229	-2,9045215

Image 43: Paste transpose to get a better view of the data

Selective export:

Use the recorder instrument to zoom into a section of your data file in <u>Dewesoft</u> (position both white cursors I and II and click between them), then the calculation and export are only done over the selected data.



Image 44: Export will be done only on the zoomed area